



U.S. Micro-Solutions, Inc. * 302 Unity Plaza * Latrobe, PA 15650
 Phone: (724) 853-4047 Fax: (724) 853-4049 AIHA-LAP, LLC EMLAP # 103009
www.usmslab.com



Customer Name:	U.S. Micro-Solutions, Inc.	Sample Date:	July 16, 2018
Customer Address:	302 Unity Plaza Latrobe, PA 15650	Date Received:	July 17, 2018
Customer Phone:	724-853-4047	Date of Report:	July 20, 2018
PO Number:		Fax:	724-853-4049
Project Name/Number:		Attention:	

Customer sample numbers below are uniquely identified by prefixing Laboratory # 12345-18

Duodenoscope Sample(s) (Bacteria) - Analytical Method USMS-B025

Sample Number	Media	Sample Description	Results of Microbial Analysis
1	SBA	Water	Total Bacterial Count <1 CFU No growth
Total Raw Count: <1 Analytical Sensitivity: 1 CFU			
2	SBA	Water	Total Bacterial Count <1 CFU No growth
Total Raw Count: <1 Analytical Sensitivity: 1 CFU			
3	SBA	Water	Total Bacterial Count 2 CFU Sphingomonas paucimobilis - 2 CFU
Total Raw Count: 2 Analytical Sensitivity: 1 CFU			

Results relate only to the samples tested. Results are reported as calculated. For biological data, the first and/or second digit should be considered significant.

When providing duplicates of this report, the document should be provided in total and not in section in accordance with AIHA-LAP, LLC. Any unauthorized or improper disclosure, copying, distribution, use, or falsification of these results is prohibited. USMS shall have no liability to the Customer or the Customer's customer for opinions stated, recommendations made, actions taken, or conduct implemented based on the test results reported.

Technical Manager: *Herbert Layman*

 Herbert Layman, BS, SM, CIEC

Sphingomonas paucimobilis - *Sphingomonas paucimobilis* is a gram-negative rod that exists in environmental niches such as water, including hospital water systems. It is not part of normal human flora.