



U.S. Micro-Solutions, Inc. * 302 Unity Plaza * Latrobe, PA 15650
 Phone: (724) 853-4047 Fax: (724) 853-4049 AIHA-LAP, LLC EMLAP #103009
www.usmslab.com

Customer Name: U.S. Micro-Solutions, Inc. Sample Date: September 25, 2020
 Customer Address: 302 Unity Plaza Date Received: November 6, 2020
 Latrobe, PA 15650 Date of Report:
 Customer Phone: (724) 853-4047 Fax: (724) 853-4049
 PO Number: Attention:
 Project Name/Number: Sample Report

Customer sample numbers below are uniquely identified by prefixing Laboratory # 12345-20

Liquid Sample(s) for Heater Cooler Unit Water Culture - Analytical Method MIC 12

Sample Number	Analysis	Sample Description	Results of Microbial Analysis
1	HPC-R2A	A	2 CFU/mL
			Raw Count: 2
AS = 1 CFU/mL (Volume examined: 2.2 mL) Analysis Time: 24 Hrs			
1	<i>Mycobacteria-7H11</i>	B	
AS = 1 CFU/mL (Volume examined: 100 mL)			Mycobacterium avium isolated

Sample Report

Notes:

Results relate only to the samples tested. Results are reported as calculated. For biological data, the first and/or second digit should be considered significant.

When providing duplicates of this report, the document should be provided in total and not in section in accordance with AIHA-LAP, LLC. Any unauthorized or improper disclosure, copying, distribution, use, or falsification of these results is prohibited. USMS shall have no liability to the Customer or the Customer's customer for opinions stated, recommendations made, actions taken, or conduct implemented based on the test results reported.

Technical Manager: Deanna L. Kiska
 Deanna L. Kiska, Ph.D.