



U.S. Micro-Solutions, Inc. * 302 Unity Plaza * Latrobe, PA 15650
 Phone: (724) 853-4047 Fax: (724) 853-4049 AIHA-LAP, LLC EMLAP # 103009
www.usmslab.com

Customer Name: U.S. Micro-Solutions, Inc. Sample Date: November 11, 2020
 Customer Address: 302 Unity Plaza Date Received: November 12, 2020
 Latrobe, PA 15650 Date of Report: November 15, 2020
 Customer Phone: (724) 853-4047 Fax: (724) 853-4049
 PO Number: Attention:
 Project Name/Number: DUO Sample Report
 Temperature upon receipt: 7.9 °C Analysis Time (hrs): 22.2
 Customer sample numbers below are uniquely identified by prefixing Laboratory # 12345-20

Duodenoscope Sample(s) (Bacteria) - Analytical Method MIC 13

Sample Number	Media	Sample Description	Results of Microbial Analysis
1	SBA	Channel flush and brushing from elevator and biopsy channel	Total Microbial Count <1 CFU No growth
Total Raw Count: <1 Analytical Sensitivity: 1 CFU			

2	SBA	Channel flush and brushing from elevator and biopsy channel	Total Microbial Count 2 CFU Staphylococcus hominis 1 CFU Staphylococcus epidermidis 1 CFU
<h2 style="text-align: center;">Sample Report</h2> Total Raw Count: 2 Analytical Sensitivity: 1 CFU			

Samples are in good condition unless otherwise noted. Results relate only to the samples tested as received. Results are reported as calculated. For biological data, the first and/or second digit should be considered significant.

When providing duplicates of this report, the document should be provided in total and not in section in accordance with AIHA-LAP, LLC. Any unauthorized or improper disclosure, copying, distribution, use, or falsification of these results is prohibited. USMS shall have no liability to the Customer or the Customer's customer for opinions stated, recommendations made, actions taken, or conduct implemented based on the test results reported.

Technical Manager:

End of Report

Deanna L. Kiska

Deanna L. Kiska, Ph.D.